

Minutes of Pre-Bid Conference (PBC) held on 04-10-2024 for proposed procurement of "Supply, installation and commissioning of "CHEMISORPTION ANALYZER WITH Q-MASS DETECTOR FOR MULTI GAS ANALYSIS"

Chairpersons / Members of the Technical Sub Committee (TSC) present during PBC including domain experts:-

1. Dr. Pratyay Basak Chairman
2. Dr. G.Jithender Reddy, Member
3. Sri. D. Venkateswara Rao, Member
4. Dr. Sreepriya Vedantam, Member
5. Dr. G.H. Gunasekar (IO)

Representatives of the following firm attended the PBC:

1. M/s Verder Scientific Pvt. Ltd., Hyderabad

The following points were discussed during the PBC:

Query raised by M/s. Verder Scientific Pvt. Ltd, and response of CSIR-IICT:

Query-1: Vapor dosing system should include a constant temperature air bath to avoid condensation of the generated vapours with appropriate temperature control.

Response: There is no change in specifications

Query-2: The Vapor dosing system should have facility to control concentration & pressure of the generated vapor using an in-built Peltier condenser or similar mechanism.

Response: There is no change in specifications

Query-3: The system should be equipped with an in-built gas mixing unit, for accurate gas mixing in any proportion arbitrarily, eliminating the need for pre-mixed gas cylinders, prepares on-line gas mixtures by using pure gas cylinders.

Response: There is no change in the specification

Query-4: It should be possible and easy for the operator to change and clean the sample tube dust filters without calling the service personnel. The system should be supplied with 20 set of filters as standard supply.

Response: The committed responded that it has already been covered in the specification.


The committee is suggested to incorporate the following points:

There is no change in the technical specifications


Points clarified by CSIR-IICT Team during PBC:


The representatives of the participating firm/further informed that they do not have any issue or suggestion with respect to other points of tendered specifications and related requirements given in the tender document. Participating bidders have been informed that points raised by them during PBC will be examined by CSIR-IICT's **Technical Sub Committee (TSC)/Technical team** constituted for the purpose of procurement of said equipment and **post PBC changes** in tendered specifications and requirements to be agreed after due consideration of the same by TSC, **if any**, will be uploaded in **CPPP** as part of **revised/amended tendered specifications** along with CSIR-IICT website www.iict.res.in on or before _____. All bidders are requested kindly to take a note of the changes, if any, in tendered specifications subsequent to **PBC** held today, i.e. 04-10-2024 before they start submitting their online bids through CPPP.


(Dr Jithender Reddy)
Member


(Dr. Sreepriya Vedantam)
Member


(Sri. D. Venkateswara Rao)
Member


(Dr. G.H. Gunasekar)
IO


(Dr. Pratyay Basak)
Chairperson
04/10/2024