

Minutes of Pre-Bid Conference (PBC) held on 29-10-2024 for proposed procurement of "Supply, installation and commissioning of "Pressure Distillaion Unit"

Chairpersons / Members of the Technical Sub Committee (TSC) present during PBC including domain experts present during PBC:-

1. Dr. N.Lingaiah, Chairman
2. Dr.PratyayBasak, Member
3. Dr.SreepriyaVedantam, Member
4. IO Mr. P.V.S.S. Srinivas

Representatives of the following firm attended the PBC:

1. M/s. PILON Engineering Pvt. Ltd.,

The following points were discussed during the PBC:


Query raised by M/s. PILON Engineering Pvt. Ltd.,and response of CSIR-IICT:


- Query-1: Does different heads required for all the three sizes of vessels
- Response: No, Single head for all the three capacities having the lifting facility
- Query-2: Whether product receivers to be jacketed? Do isolation valves are required
- Response: Yes, all the product receivers are to be provided with jacketed and also isolation valves
- Query-3: Type of warranty required
- Response: Two years of comprehensive


K. Mahesh


**Points clarified by CSIR-IICT Team during PBC:**

The representatives of the participating firm/further informed that they do not have any issue or suggestion with respect to other points of tendered specifications and related requirements given in the tender document. Participating bidders have been informed that points raised by them during PBC will be examined by CSIR-IICT's **Technical Sub Committee (TSC)/Technical team** constituted for the purpose of procurement of said equipment and **post PBC changes** in tendered specifications and requirements to be agreed after due consideration of the same by TSC, **if any**, will be uploaded in **CPPP** as part of **revised/amended tendered specifications** along with CSIR-IICT website [www.iict.res.in](http://www.iict.res.in) on or before **25-09-2024**. All bidders are requested kindly to take a note of the changes, if any, in tendered specifications subsequent to **PBC** held today, i.e. 29-10-2024 before they start submitting their online bids through CPPP.

  
(Dr. Pratyay Basak)  
Member

  
(Dr. Sreepriya Vedantam)  
Member


  
(Mr. P. V. S. S. Srinivas)  
IO


  
(Dr. N. Lingaiah)  
Chairperson


The following changes has been made in tendered specification subsequent to PBC held on 29.10.2024

S. No.	Existing Specifications	Revised/Amended Specifications
1	Reboiler: 2I,3I & 5I (Interchangeable)	Single head for all the three capacities having the lifting facility
2.	Product Receivers: Jacketed	Jacketed vessels with isolation valves
3.	Warranty: 2 years	2 years of comprehensive warranty


All the other tender terms remain unchanged. Bidders may please submit their bids accordingly.

  
(Dr. Pratyay Basak)  
Member

  
Dr. Sreepriya Vedantam)  
Member

  
(Mr. P.V.S.S. Srinivas)  
IQ

  
(Dr. L. Ravitej Singh)  
PL

  
(Dr N Lingaiah)  
Chairperson