

CPP TENDER ID: 2024\_CSIR\_208132\_1 Dtd 27.09.2024

PUR/IICT/1011/24-25/EQPT

CORRIGENDUM

**Supply, Installation and commissioning of "CUSTOMIZED CONTINUOUS REACTIVE DISTILLATION COLUMN**


The following corrigendum is issued to above mentioned tender for supply of "Customized Continuous Reactive Distillation Column"

3(b) Heating Circulator

**Tendered Specifications: Temperature Stability +/- 0.001**

**Revised Technical Specification: Temperature Stability +/- 0.01**

Bidders may please take note of the above changes and submit their quotations accordingly. All other technical specifications including due date of bid submission remain unchanged.

A handwritten signature in blue ink, followed by the date "27.09.24" written in blue ink.

Controller of Stores & Purchase

CSIR-IICT