



## CORRIGENDUM

No. PUR/IICT/0922/24-25/EQPT

Date:10.09.2024

**Sub: Corrigendum/Clarification on the tendered Specifications for procurement of "Dosing Pump"-reg.**

**Ref: CPPP Id No. 2024\_CSIR\_206938\_1 dt. 03-09-2024.**

The aforesaid tender under reference for proposed procurement of "**Dosing Pump**" the following amendment/clarifications are issued for information and reference of all the potential bidders-

S.No.	Points raised by the bidder	Amendment/ Clarification on existing specifications
1	Pump Type (Plunger Type/ Diaphragm Type/ Electronic diaphragm type) ?	<b>Diaphragm Type</b>
2	Name of Liquid (Viscosity of liquid/ Density of liquid) ?	<b>Usage for different types of liquids, viscosity: <math>\geq 150</math> cSt, Density : <math>\leq 1.5</math>g/cc</b>
7	Flowrate in LPH ?	<b>Min. <math>\geq 3</math>ml/hr &amp; Max <math>\geq 1200</math>ml</b>
8	Pressure in Kg/cm <sup>2</sup> ?	<b><math>\geq 6</math> bar/Kg/Cm<sup>2</sup></b>
10	Connections (BSP/ Flanged) ?	<b>BSP Fitting</b>
11	Motor (Flameproof/ non-flameproof)?	<b>Flame proof</b>
15	Accessories (Strainer/ External PRV/ Discharge dampener)?	<b>Strainer, External PRV required</b>

All the bidders are requested to kindly take a note of aforesaid clarifications and submit the quote accordingly, which will be considered as part of tendered specifications for evaluation of technical bids submitted by the bidders.

All other terms & conditions of the tendered document remain unchanged.

  
(Dharmendra Kumar)  
Controller of Stores & Purchase