

File Ref. No. PUR/IICT/DMS/1068/RE/23-24
CPPP Tender ID : 2024_CSIR_202965_1

Dt: 06-08-2024

Minutes of Pre-Bid Conference (PBC) held on 06-08-2024 for proposed procurement of "Supply, installation and commissioning of "GEL PERMEATION CHROMOTOGRAPHY - 1 No."

Chairpersons / Members of the Technical Sub Committee (TSC) present during PBC including domain experts present during PBC:-

1. Dr. N.Lingaiah, Chairman
2. Dr.G.Jithender Reddy, Member
3. Dr.SreepriyaVedantam, Member
4. IO Dr. P. Aruna
5. Dr. Arindam Chakrabarty – PFM Dept.
6. Dr. K.S.N. Raju - PFM Dept.

Representatives of the following firm attended the PBC:

1. M/s CAMTEK LABS

The following points were discussed during the PBC:

Query raised by M/s. CAMTEK LABS, and response of CSIR-IICT:

Query-1:RID cell volume to be modified from 10-20 μ L to 8-20 μ L

Response: Agreed. The specifications of RID cell volume will be amended to 8-20 μ L


Points clarified by CSIR-IICT Team during PBC:

The representatives of the participating firm/further informed that they do not have any issue or suggestion with respect to other points of tendered specifications and related requirements given in the tender document. Participating bidders have been informed that points raised by them during PBC will be examined by CSIR-IICT's **Technical Sub Committee (TSC)/Technical team** constituted for the purpose of procurement of said equipment and **post PBC changes** in tendered specifications and requirements to be agreed after due consideration of the same by TSC,

if any, will be uploaded in **CPPP** as part of **revised/amended tendered specifications** along with CSIR-IICT website www.iict.res.in or before **09-08-2024**. All bidders are requested kindly to take a note of the changes, if any, in tendered specifications subsequent to **PBC** held today, i.e. **06-08-2024** before they start submitting their online bids through CPPP.

 06/08/24

(Dr Jithender Reddy)
Member

 06.08.2024

(Dr. Sreepriya Vedantam)
Member



(Dr. P. Aruna)
IO

6/8/24



(Dr. Rohit Kumar Rana)
PL



(Dr N Lingaiah) 6/8/2024
Chairperson

Revised Specifications/Corrigendum

File Ref. No. PUR/IICT/DMS/1068/RE/23-24

Dt: 06-08-2024

CPPP Tender ID : 2024_CSIR_202965_1


The following changes has been made in tendered specification subsequent to PBC held on 06.08.2024


S. No.	Existing Specifications	Revised/Amended Specifications
1	RID cell volume 10-20 µL.	RID cell volume will be amended to 8-20 µL.


All the other tender terms remains unchanged. Bidders may please submit their bids accordingly.


06/08/24
Dr Jithender Reddy)
Member


06.08.2024
(Dr. Sreepriya Vedantam)
Member


(Dr. P. Aruna)
IO
6/8/24


(Dr. Rohit Kumar Rana)
PL


(Dr N Lingaiah) 6/8/2024
Chairperson