

Minutes of Pre-Bid Conference (PBC) held on 06-08-2024 for proposed procurement of "Supply, installation and commissioning of "GC WITH FID AND NPD DETECTORS - 1 No."

Chairpersons / Members of the Technical Sub Committee (TSC) present during PBC including domain experts present during PBC:-

1. Dr. N.Lingaiah, Chairman
2. Dr.G.Jithender Reddy, Member
3. Dr.SreepriyaVedantam, Member
4. IO Dr. G.V. Karunakar
5. 5. G. Balakrishna - FAC Division

Representatives of the following firm attended the PBC:

1. M/s. Smart Labtech Pvt. Ltd.,
2. M/s. Camtek Labs, Hyderabad

The following points were discussed during the PBC:

Query raised by M/s.Camtek Labs, Hyderabad and response of CSIR-IICT:

Query-1: Whether Additional Software with License is needed to be installed in other 1 PC
Response: Yes, License with software to be installed in 1 more PC for offline data process.


Query raised by M/s. Smart Labtech Pvt Ltd, Hyd , and response of CSIR-IICT:

Query-1: Asked to amend the Set data rate from 1000 to 600 Hz
Response: Yes, Accepted.


Points clarified by CSIR-IICT Team during PBC:

The representatives of the participating firm/further informed that they do not have any issue or suggestion with respect to other points of tendered specifications and related requirements given in the tender document. Participating bidders have been informed that points raised by them during PBC will be examined by CSIR-IICT's **Technical Sub Committee (TSC)/Technical team** constituted for the purpose of procurement of said equipment and **post PBC changes** in tendered specifications and requirements to be agreed after due consideration of the same by TSC, **if any**, will be uploaded in CPPP as part of **revised/amended tendered specifications** along with CSIR-IICT website www.iict.res.in on or before **09.08.2024** . All bidders are requested kindly to take a note of the changes, if any, in tendered specifications subsequent to **PBC** held today, i.e. 06-08-2024 before they start submitting their online bids through CPPP.


(Dr Jithender Reddy)
Member


(Dr. G.V. Karunakar)
I/o


(Dr.Sreepriya Vedantam)
Member


(Dr. A Krishnaiah)
IO /PL


(Dr N Lingaiah)
Chairperson

Revised Specifications/Corrigendum

File Ref. No. PUR/IICT/DMS/0549/24-25
CPPP Tender ID : 2024_CSIR_203043_1

Dt: 06-08-2024


The following changes has been made in tendered specification subsequent to PBC held on 06.08.2024


S. No.	Existing Specifications	Revised/Amended Specifications
1	FID Must be able to set data rate of 1000 Hz or more	FID Must be able to set data rate of 600 Hz or more
2	Should provide offline data processing software for multiple Pcs.	License with software to be installed in one more PC for offline data process.


All the other tender terms remains unchanged. Bidders may please submit their bids accordingly.


Dr Jithender Reddy
Member


(Dr. Sreepriya Vedantam)
Member


(Dr. G V Karunakar)
IO


(Dr. A Krishnaiah)
IO/PL


(Dr N Lingaiah)
Chairperson