

**Minutes of Pre-Bid Conference (PBC) held on 04-09-2024 for proposed procurement of "HPLC - 1 No."**

**Chairpersons / Members of the Technical Sub Committee (TSC) present during PBC including domain experts present during PBC:-**

1. Dr. N. Lingaiah, Chairman
2. Dr. Pratyay Basak, Member
3. Dr. G. Jithender Reddy, Member
4. Sri. D. Venkateswara Rao, Member
5. Dr. Sreepriya Vedantam, Member
6. IO Dr. Sandip B. Bharate

**Representatives of the following firm attended the PBC:**

1. M/S Thermo Fisher Scientific Pvt Ltd
2. M/S Spinco Biotech Pvt Ltd
3. M/S Waters India Pvt. Ltd.

**The following points were discussed during the PBC:**

**Query raised by M/s. Spinco Biotech Pvt Ltd, and response of CSIR-IICT:**

**Query-1:** Point no. 1 of technical specifications. Requested to change the gradient accuracy to +/- 1%.

**Response:** Agreed.

**Query-2:** Point no. 1 of technical specifications. Requested to change the gradient precision to +/- 0.5%.

**Response:** Agreed.


**Additional change made by the CSIR-IICT, which was also discussed during the PBC.**

**Point no. 13:** "Suitable PC & Printer" is replaced with "PC with i7 processor, 8 GB RAM and 1TB SSD along with autoduplex black & white LaserJet printer"


**Points clarified by CSIR-IICT Team during PBC:**




The representatives of the participating firm/further informed that they do not have any issue or suggestion with respect to other points of tendered specifications and related requirements given in the tender document. Participating bidders have been informed that points raised by them during PBC will be examined by CSIR-IICT's **Technical Sub Committee (TSC)/Technical team** constituted for the purpose of procurement of said equipment and **post PBC changes** in tendered specifications and requirements to be agreed after due consideration of the same by TSC, **if any**, will be uploaded in **CPPP** as part of **revised/amended tendered specifications** along with CSIR-IICT website [www.iict.res.in](http://www.iict.res.in) on or before **06-09-2024**. All bidders are requested kindly to take a note of the changes, if any, in tendered specifications subsequent to **PBC** held today, i.e. 04-09-2024 before they start submitting their online bids through CPPP.


  
05/09/2024  
(Dr. Pratyay Basak)  
Member

  
(Dr. Jithender Reddy)  
Member

  
(Dr. Sreepriya Vedantam)  
Member

  
(Sri. D. Venkateswara Rao)  
Member


  
04/09/2024  
(Dr. Sandip B. Bharate)  
IO/PL

  
5/9/2024  
(Dr. N. Lingaiah)  
Chairman


The following changes has been made in tendered specification subsequent to PBC held on 04.09.2024

S. No.	Existing Specifications	Revised/Amended Specifications
1	<ul style="list-style-type: none"><li>• Gradient accuracy: +/- 0.5% independent of back pressure or less</li><li>• Gradient precision: +/- 0.15% independent of back pressure or less</li></ul>	<ul style="list-style-type: none"><li>• Gradient accuracy: +/- 1% independent of back pressure or less</li><li>• Gradient precision: +/- 0.5% independent of back pressure or less</li></ul>
13	<ul style="list-style-type: none"><li>• Suitable PC &amp; Printer</li></ul>	<ul style="list-style-type: none"><li>• PC with i7 processor, 8 GB RAM and 1TB SSD along with autoduplex black &amp; white LaserJet printer</li></ul>


All the other tender terms remain unchanged. Bidders may please submit their bids accordingly.


  
(Dr. Pratyay Basak)  
Member

  
(Dr. Jithender Reddy)  
Member

  
(Dr. Sreepriya Vedantam)  
Member

  
(Sri. D. Venkateswara Rao)  
Member

  
(Dr. Sandip B. Bharate)  
IO/PL

  
(Dr. N Lingaiah)  
Chairman